



## **Appendix 2**

**This appendix was part of the submitted manuscript and has been peer reviewed. It is posted as supplied by the authors.**

Appendix to: Wong M, Lim C, Lee K-L, et al. An unusual case of implantable cardioverter-defibrillator inhibition. *Med J Aust* 2015; 202: 329. doi: 10.5694/mja14.01166.

Appendix 2: Individual contribution of quantitative magnetic field strengths showing the vertical (z axis, in red) contributing to the highest field strength (800 Gauss) inhibiting the bra magnet

